


CPS® Products

This is to certify that CPS model ABM-HOOD airflow, temperature and humidity instrument is calibrated to NIST traceable standards using one or more of the instruments listed below:

Omega HH42A
Setra 370
Digi-Sense 20250-11
Omega WTM-1000
Testo 622
WindSpeed Tunnel JIG

This ABM-HOOD using instruments ABM-100 or ABM200 for airflow sensing was tested and confirmed for published accuracy per ANSI/AMCA 210-07 Laboratory Methods of testing Fans for Certified Aerodynamic Performance Rating.



Vice President of Engineering
CPS Products, Inc.

August 23, 2017

Date.

CPS WORLD HEADQUARTERS

1010 East 31st Street,
Hialeah, Florida 33013, USA
Tel: 305-687-4121, 1-800-277-3808
Fax: 305-687-3743
E-mail: info@cpsproducts.com
Website: www.cpsproducts.com